



**Sea-Bird Electronics, Inc.**  
 13431 NE 20th Street  
 Bellevue, WA 98005 United States

Phone  
 Fax

+1-425-643-9866  
 +1-425-643-9954  
 www.seabird.com

**SERVICE REPORT**

**Service Request**  
**Date**  
**Sales Order**

**1005502977**  
 06-DEC-2017  
 314828000

**CUSTOMER INFORMATION**

Name: KONGSBERG UNDERWATER TECHNOLOGY INC  
 Account : 40280537  
 RICHARD HILE  
 richard.hile@kongsberg.com  
 (425) 245-0557

PO Number:  
 PO 01323G

**Bill To Address**

19210 33RD AVE W;SUITE A;  
 LYNNWOOD,WA,98036,US

**Ship To Address**

19210 33RD AVE W;SUITE A;  
 LYNNWOOD,WA,98036,US

**PRODUCT INFORMATION**

**Item:** APL-GLIDER.LEGACY  
**Item Description:** (LEGACY) APL Glider  
**Serial:** 0073

**Special Notes**

Services Requested:  
 Evaluate/Repair Instrumentation.  
 Perform Routine Calibration Service.

**Problems Found:**

The conductivity board was found to have failed and will not pass calibration.

**Services Performed:**

Perform initial diagnostic evaluation.  
 Replaced conductivity cell.  
 Replaced conductivity board.  
 Performed a "Final" Calibration.

Item	Item Description	Qty
41119T	UW/APL GLIDER ISOLATED WBO COND ELECT, TESTED	1
CAL_SEAGLIDER	Calibrate Seaglider conductivity and temperature sensors	1
CNCRTSEAGLDR	Confirm & Re-certify Webb SLOCUM Glider CTD	1

**Unbilled Items**

Item	Item Description	Qty
801912	SBE 41 ALACE COND CELL ASSEMBLY, NO WRAP	1



**Sea-Bird Electronics, Inc.**  
 13431 NE 20th Street  
 Bellevue, WA 98005 United States

Phone  
 Fax

+1-425-643-9866  
 +1-425-643-9954  
 www.seabird.com

**SERVICE REPORT**

**Service Request**  
**Date**  
**Sales Order**

**1005502977**  
 06-DEC-2017  
 314828000

**PRODUCT INFORMATION**

**Item:** 43F.LEGACY  
**Item Description:** (LEGACY) SBE 43F Oxygen Sensor  
**Serial:** 43F0221

**Special Notes**  
 Services Requested:  
 Evaluate/Repair Instrumentation.  
 Perform Routine Calibration Service.

Item	Item Description	Qty
CAL_43	SBE 43 dissolved oxygen sensor calibration	1

**Unbilled Items**

Item	Item Description	Qty