≮ISTLER

measure. analyze. innovate.

Piezoresistive Pressure Sensor Calibration

Туре			Certificate ID #			4813657-151229T1528					
Serial Number	4813657 Kistler e 0 to 1500 PSI Absolute				Calibration Technician Date/Time Snan			Chris Prell 12/29/2015 3:28:55 PM			
Manufacturer											
Pressure Range								nV/V	9.896		
Reference					Offset	n	nV/V	1.076			
Tost Condition New					Sunnly Voltago			/	9 995		
					Sobbia AO	inage	•		0.000		
Non-Linearity, H	ysteresis, and Re	peatability (N	LHR)		Summary:	:					
P (PSI)	Output (mV/V)	BFSL Error (9	%)								
0.0	1.0756	0.015									
375.0	3.5471	-0.010									
750.0	6.0207	-0.013									
1125.0	8.4956	-0.002									
1500.0	10.9712	0.011									
750.0*	6.0210	0.003									
0.0*	1.0756	0.000			Envrionme	ental Conditio	ns				
* Decreasing Pressure					Tempera	ature		°C	22 ± 4		
					Relative Humidity			%	30 ± 30		
Temperature Per	rformance										
Temperature (°C)	Span Error (%)	Offset Error (%)								
-8.7	-0.166	-0.024			0.15						
-0.1	-0.132	-0.026			0.1						
25.0	0.000	0.000			0.1					*******	
40.1	0.115	0.044			0.05					••	
				-	0				and a second		
Error Calculation Unit				%	-0.05						
NLHR limits are based on			% span		0.00			******			
Temperature Performance limits are based on			% span		-0.1		and a state of the				
Span & Offset limits are based on % sp			% span		-0.15						
					-0.2	-			1		
					-20		0	2	0	40	
							Te	mperature	(°C)		
								-			
							· Span ··	•••• Offset			
							-				
Reference Equip	ment										
Туре	S/N										

This sensor was calibrated per Kistler test procedure 680-0000-701 using a comparison technique against a Kistler working standard. Kistler working standards are periodically calibrated against a primary standard system, which in turn is periodically recertified to the National Institute of Standards and Technology (NIST) or another recognized national standard. Measurements are derived from accepted values of natural physical constants according to the International System of Units (SI). This calibration meets or exceeds the requirements of MIL-STD-45662A, ISO 9001, ANSI/NCSL Z540-1 and is accredited to ISO/IEC 17025 as verified by the ANSI-ASQ National Accreditation Board/ACLASS. Refer to certificate and Scope of Accreditation AC-1117. Estimated uncertainty of this calibration is $\pm 0.2\%$ of pressure range for voltage output sensors or $\pm 0.25\%$ of pressure range for current output sensors with respect to the primary standard. Certificates are on file at Kistler and may be requested in writing. This certificate shall not be reproduced, except in full, without written approval of Kistler Instrument Corporation.

Kistler Instrument Corp. 75 John Glenn Drive Amherst, NY 14228-2171 Tel 1-888-KISTLER Fax 1-716-691-5226 info.us@kistler.com ISO 9001 CERTIFIED QUALITY SYSTEM ISO 17025 Accredited Calibration Laboratory 040-0096-001 Rev D Page 2 of 2 www.kistler.com