



CALIBRATION CERTIFICATE

Form No. 622, Dec 2005

a xylem brand

Sensing Foil Batch No: 1023
Certificate No: 5013 19 1268

Product: 5013
Serial No: 19
Calibration Date: September 12, 2014

This is to certify that this product has been calibrated using the following instruments:

Fluke CHUB E-4	Serial No. A7C677
Fluke 5615 PRT	Serial No. 849155
Fluke 5615 PRT	Serial No. 802054
Honeywell PPT	Serial No. 44074
Calibration Bath model FNT 321-1-40	1

Parameter: Internal Temperature:

Calibration points and readings:

Temperature (°C)	-	-	-	-
Reading (mV)	-	-	-	-

Giving these coefficients

Index	0	1	2	3
TempCoef	2.65834E+01	-3.10940E-02	2.93876E-06	-4.33055E-09

*Note: Temperature calibration NOT performed

Parameter: Oxygen:

	O2 Concentration	Air Saturation
Range:	0-500 μM ¹⁾	0 - 120%
Accuracy ¹⁾ :	< $\pm 8\mu\text{M}$ or $\pm 5\%$ (whichever is greater)	$\pm 5\%$
Resolution:	< 1 μM	< 0.4%
Settling Time (63%):	< 25 seconds	

Calibration points and readings²⁾:

	Air Saturated Water	Zero Solution (Na ₂ SO ₃)
Phase reading (°)	3.35490E+01	6.50264E+01
Temperature reading (°C)	9.99469E+00	2.16274E+01
Air Pressure (hPa)	1.01300E+03	

Giving these coefficients

Index	0	1	2	3
PhaseCoef	-7.40472E+00	1.22176E+00	0.00000E+00	0.00000E+00

¹⁾ Valid for 0 to 2000m (6562ft) depth, salinity 33 - 37ppt

²⁾ The calibration is performed in fresh water and the salinity setting is set to: 0

Date:
September 12, 2014

Signed: Jason B Baxter

Aanderaa Data Instruments

1700/1725 Brannum Lane | Yellow Springs, OH 45387 | Tel. +1-937-767-4974 | email: infoUSA@xyleminc.com