



# CALIBRATION CERTIFICATE

Form No. 622, Dec 2005

a xylem brand

Sensing Foil Batch No: 1707  
Certificate No: 5013 21 1115

Product: 5013  
Serial No: 21  
Calibration Date: November 29, 2011

This is to certify that this product has been calibrated using the following instruments:

Fluke CHUB E-4	Serial No. A7C677
Fluke 5615 PRT	Serial No. 849155
Fluke 5615 PRT	Serial No. 802054
Honeywell PPT	Serial No. 44074
Calibration Bath model FNT 321-1-40	1

Parameter: Internal Temperature:

Calibration points and readings:

Temperature (°C)	-	-	-	-
Reading (mV)	-	-	-	-

Giving these coefficients

Index	0	1	2	3
TempCoef	2.43177E+01	-3.11208E-02	2.93847E-06	-4.22055E-09

\*Note: Temperature calibration not performed

Parameter: Oxygen:

	O2 Concentration	Air Saturation
Range:	0-500 $\mu\text{M}$ <sup>1)</sup>	0 - 120%
Accuracy <sup>1)</sup> :	< $\pm 8\mu\text{M}$ or $\pm 5\%$ (whichever is greater)	$\pm 5\%$
Resolution:	< 1 $\mu\text{M}$	< 0.4%
Settling Time (63%):	< 25 seconds	

Calibration points and readings<sup>2)</sup>:

	Air Saturated Water	Zero Solution ( $\text{Na}_2\text{SO}_3$ )
Phase reading (°)	3.27612E+01	6.44009E+01
Temperature reading (°C)	1.00170E+01	2.24060E+01
Air Pressure (hPa)	9.84300E+02	

Giving these coefficients

Index	0	1	2	3
PhaseCoef	-2.42427E+00	1.17630E+00	0.00000E+00	0.00000E+00

<sup>1)</sup> Valid for 0 to 2000m (6562ft) depth, salinity 33 - 37ppt

<sup>2)</sup> The calibration is performed in fresh water and the salinity setting is set to: 0

Date:  
November 29, 2011

Sign: Shawn A. Sneddon

Service and Calibration Engineer

Aanderaa Data Instruments, Inc.

182 East Street, Suite B Attleboro, MA 02703 Tel. +1 (508) 226-9300 email: infoUSA@xyleminc.com





# CALIBRATION CERTIFICATE

Form No. 622, Dec 2005


a xylem brand

Sensing Foil Batch No: 1707  
Certificate No: 5013 21 1115

Product: 5013  
Serial No: 21  
Calibration Date: November 29, 2011

## Data from Cool Down Test:

### Cool Down Test



## SR10 Scaling Coefficients:

At the SR10 output the Oxygen Optode 3830 can give either absolute oxygen concentration in  $\mu\text{M}$  or air saturation in %. The setting of the internal property "Output"<sup>3)</sup>, controls the selection of the unit. The coefficients for converting SR10 raw data to engineering units are fixed.

Output = -1	Output = -2
A = 0	A = 0
B = 4.883E-01	B = 1.465E-01
C = 0	C = 0
D = 0	D = 0
Oxygen ( $\mu\text{M}$ ) = A + BN + CN2 + DN3	Oxygen (%) = A + BN + CN2 + DN3

<sup>3)</sup> The default output setting is set to -1

Date:  
November 29, 2011

Sign: Shawn A. Sneddon

Service and Calibration Engineer

Aanderaa Data Instruments, Inc.

182 East Street, Suite B Attleboro, MA 02703 Tel. +1 (508) 226-9300 email: infoUSA@xyleminc.com



# TEST & SPECIFICATIONS

Form No. 620, Nov 2005

a xylem brand

Layout No:  
Circuit Diagram No:  
Program Version:

Product: 5013  
Serial No: 21

- 
- |  |           |
|--|-----------|
| <b>1. Visual and Mechanical Checks:</b>                  |           |
| 1.1. O-ring surface                                      | N/A       |
| 1.2. Soldering quality                                   | N/A       |
| 1.3. Visual surface                                      | OK        |
| 1.4. Pressure test (60MPa)                               | N/A       |
| 1.5. Galvanic isolation between housing and electronics  | OK        |
| <b>2. Current Drain and Voltages:</b>                    |           |
| 2.1. Average current drain at 0.5Hz sampling (Max: 38mA) | 31.4 mA   |
| 2.2. Current drain in sleep (Max: 300uA)                 | 190 uA    |
| <b>3. Performance Test in Air, 20°C Temperature:</b>     |           |
| 3.1. Amplitude measurement (Blue: 290 – 470mV)           | 378.63 mV |
| 3.2. Phase measurement (Blue: 27 ±5°)                    | 29.1 °    |
| 3.3. Temperature Measurement (100 ± 300mV)               | -52.28 mV |
| <b>4. Firmware:</b>                                      |           |
| 4.1. Firmware upgrade                                    | 3.24      |

Date:  
November 29, 2011

Sign: Shawn A. Sneddon

Service and Calibration Engineer

Aanderaa Data Instruments, Inc.

182 East Street, Suite B    Attleboro, MA 02703    Tel. +1 (508) 226-9300    email: [infoUSA@xyleminc.com](mailto:infoUSA@xyleminc.com)





# CALIBRATION CERTIFICATE

Form No. 621, Dec 2005

a xylem brand

Sensing Foil Batch No: 1707

Certificate No:

Product: O2 Sensing Foil PSt3 3853

Calibration Date: 22 June 2007

## Calibration points and phase readings (degrees)

Temperature (°C)		4.54	10.98	20.44	29.92	39.26
Pressure (hPa)		968.25	968.25	968.25	968.25	968.25
O2 in % of O2+N2	0.00	74.37	73.85	73.04	72.17	70.99
	1.00	69.53	68.59	67.18	65.72	64.11
	2.00	66.04	64.85	63.10	61.34	59.52
	5.00	57.77	56.17	53.91	51.75	49.74
	10.00	48.62	46.85	44.43	42.22	40.29
	20.90	37.69	36.01	33.81	31.89	30.28
	30.00	32.57	31.02	29.05	27.36	25.96

Giving these coefficients <sup>1)</sup>

Index	0	1	2	3
C0 Coefficient	5.32650E+03	-1.92117E+02	4.14357E+00	-3.78695E-02
C1 Coefficient	-2.92068E+02	9.71993E+00	-2.14295E-01	2.00778E-03
C2 Coefficient	6.47595E+00	-1.98080E-01	4.49940E-03	-4.30530E-05
C3 Coefficient	-6.69288E-02	1.88066E-03	-4.42348E-05	4.28382E-07
C4 Coefficient	2.65042E-04	-6.83185E-06	1.67071E-07	-1.61989E-09

<sup>1)</sup> Ask for Form No 621S when this O2 Sensing Foil is used in Oxygen Sensor 3830 with Serial Numbers lower than 184.

Date:

June 22, 2007

Aanderaa Data Instruments, Inc.

182 East Street, Suite B

Attleboro, MA 02703

Tel. +1 (508) 226-9300

email: [infoUSA@xyleminc.com](mailto:infoUSA@xyleminc.com)



DATE: November 8, 2011  
Prepared by Shawn Sneddon  
Service Order 2584

**AANDERAA DATA INSTRUMENTS**  
**US SERVICE & CALIBRATION DEPARTMENT**

**Service Report**

**Oxygen Optode 5013 sn21**

1. Performed visual inspection
  - a. Slight separation between cable and potting
    - i. Recommend re-splicing and potting with new mold.
2. Checked for isolation between housing and electronics
  - a. Isolation OK.
3. Checked current consumption
  - a. Operating = 31.4mA; OK.
  - b. Quiescent = 190uA; OK.
4. Performed test in air checking BAmp, BPhase, and RawTemp
  - a. All OK.
5. Inspected foil visually
  - a. Looks OK.
6. Checked firmware version
  - a. 3.22; OLD.
  - b. Upgraded to version 3.24.
7. Checked temperature in 10 deg.C bath with reference
  - a. Sn21 = 9.98, Reference = 9.976; OK.
8. Checked saturation in 100% saturated bath with reference optode
  - a. Sn21 = 98.36%, sn338 = 96.500%; Within spec
    - i. Customer requested recalibration.
9. Optode opened to repair cable
  - a. Removed potting and re-spliced cable.
  - b. Tested output: OK.
  - c. Potted cable.
  - d. Tested output: OK.
  - e. Tested current consumption.
    - i. Operating = mA; .
    - ii. Quiescent = uA; .
  - f. O-ring 3025 replaced.
  - g. BPot value set to 4 to allow optimal BAmp value.
10. Performed saturation calibration at 100% and 0% saturation
  - a. PASSED.
11. Checked saturation in 100% saturated bath with reference optode

- a. Sn21 = 97.28%, sn338 = 97.229%; OK.
- 12. Checked saturation in 20 deg.C bath with reference optode
  - b. Sn21 = 96.98%, sn338 = 97.903%; OK.
- 13. Performed cool down test from 20 to 1 deg.C
  - a. PASSED.
- 14. Returned to customer settings

**Oxygen Optode 4330IE sn199**

- 1. Performed visual inspection
  - a. OK.
- 2. Checked for isolation between housing and electronics
  - a. Isolation OK.
- 3. Checked current consumption
  - a. Operating = 30.6; OK.
  - b. Quiescent = 115uA; OK.
- 4. Performed test in air checking C1Amp, C2Amp, TCPhase, and RawTemp
  - a. C1Amp slightly low; all others OK.
- 5. Inspected foil visually
  - a. Looks slightly bleach; may need to be replaced.
- 6. Checked firmware version
  - a. 1.22.1; OLD.
  - b. Could not upgrade; boot enable not pinned out.
- 7. Checked temperature in 10 deg.C bath with reference
  - a. Sn199 = 9.987, Reference = 9.976; OK.
- 8. Checked saturation in 100% saturated bath with reference optode
  - a. Sn199 = 73.637%, sn338 = 96.500%; Needs to be recalibrated.
- 9. Performed saturation calibration at 100% and 0% saturation
  - a. FAILED two calibrations; foil should be replaced.
- 10. Replaced foil with batch 1023E and burned in overnight
  - a. Adjusted gain for optimal C1Amp and C2Amp values.
- 11. Checked saturation in 100% saturated bath with reference optode
  - a. Sn199 = 95.638%, sn338 = 96.277%; OK.
- 12. Checked saturation in 20 deg.C bath with reference optode
  - a. Sn199 = 94.955%, sn338 = 95.608%; OK.
- 13. Performed cool down test from 20 to 1 deg.C
  - a. PASSED.
- 14. Returned to customer settings

Next Calibration Date: November 29, 2013

Next Service Date: November 29, 2013