

PO Box 518
620 Applegate St.
Philomath OR 97370
wetlabs@wetlabs.com



(541) 929-5650
Fax (541) 929-5277
www.wetlabs.com

2/16/15
Date: ~~4/3/14~~

Customer: Univ of Hawaii/SOEST

S/N# BBFL2VMT-402

Technician: dcm

Repairs and Modifications: Evaluated instrument and found instrument fully functional, conducted standard puck service

Comments: New char sheets included.

PO Box 518
620 Applegate St.
Philomath, OR 97370



(541) 929-5650
Fax (541) 929-5277
www.wetlabs.com

Scattering Meter Calibration Sheet

2/16/2015

Wavelength: 650

S/N BBFL2VMT-402

Use the following equation to obtain "scaled" output values:

$$\beta(\theta_c) \text{ m}^{-1} \text{ sr}^{-1} = \text{Scale Factor} \times (\text{Output} - \text{Dark Counts})$$

- **Scale Factor for 650 nm** = 4.085E-06 (m⁻¹sr⁻¹)/counts
- **Output** = meter reading counts
- **Dark Counts** = 50 counts

Instrument Resolution = 1.5 counts 6.02E-06 (m⁻¹sr⁻¹)

Definitions:

- **Scale Factor:** Calibration scale factor, $\beta(\theta_c)/\text{counts}$. Refer to User's Guide for derivation.
- **Output:** Measured signal output of the scattering meter.
- **Dark Counts:** Signal obtained by covering detector with black tape and submersing sensor in water.

Instrument Resolution: Standard deviation of 1 minute of collected data.

PO Box 518
620 Applegate St.
Philomath, OR 97370



(541) 929-5650
Fax (541) 929-5277
www.wetlabs.com

ECO CDOM Fluorometer Characterization Sheet

Date: 2/16/2015

S/N: BBFL2VMT-402

CDOM concentration expressed in ppb can be derived using the equation:

$$\text{CDOM (ppb)} = \text{Scale Factor} * (\text{Output} - \text{Dark Counts})$$

Dark Counts

Scale Factor (SF)

Maximum Output

Resolution

Digital

45 counts

0.1335 ppb/count

4124 counts

1.3 counts

Ambient temperature during characterization

21.0 °C

Dark Counts: Signal output of the meter in clean water with black tape over detector.

SF: Determined using the following equation: $SF = x \div (\text{output} - \text{dark counts})$, where x is the concentration of the solution used during instrument characterization. SF is used to derive instrument output concentration from the raw signal output of the fluorometer.

Maximum Output: Maximum signal output the fluorometer is capable of.

Resolution: Standard deviation of 1 minute of collected data.

PO Box 518
620 Applegate St.
Philomath, OR 97370



(541) 929-5650
Fax (541) 929-5277
www.wetlabs.com

ECO Chlorophyll Fluorometer Characterization Sheet

Date: 2/16/2015

S/N: BBFL2VMT-402

Chlorophyll concentration expressed in $\mu\text{g/l}$ can be derived using the equation:

$$\text{CHL } (\mu\text{g/l}) = \text{Scale Factor} * (\text{Output} - \text{Dark counts})$$

Dark counts	Digital 46 counts
Scale Factor (SF)	0.0112 $\mu\text{g/l/count}$
Maximum Output	4130 counts
Resolution	1.0 counts
Ambient temperature during characterization	21.0 $^{\circ}\text{C}$

Dark Counts: Signal output of the meter in clean water with black tape over detector.

SF: Determined using the following equation: $\text{SF} = x \div (\text{output} - \text{dark counts})$, where x is the concentration of the solution used during instrument characterization. SF is used to derive instrument output concentration from the raw signal output of the fluorometer.

Maximum Output: Maximum signal output the fluorometer is capable of.

Resolution: Standard deviation of 1 minute of collected data.

The relationship between fluorescence and chlorophyll-a concentrations in-situ is highly variable. The scale factor listed on this document was determined using a mono-culture of phytoplankton (*Thalassiosira weissflogii*). The population was assumed to be reasonably healthy and the concentration was determined by using the absorption method. To accurately determine chlorophyll concentration using a fluorometer, you must perform secondary measurements on the populations of interest. This is typically done using extraction-based measurement techniques on discrete samples. For additional information on determining chlorophyll concentration see "Standard Methods for the Examination of Water and Wastewater" part 10200 H, published jointly by the American Public Health Association, American Water Works Association, and the Water Environment Federation.



Instrument Checklist

Date: 2/17/2015

S/N: BBFL2VMT-402

Order # 26016

Contents:

Description	Qty	Packed
■ ECO Meter	1	X
■ Calibration/Characterization Sheet	1	X
■ Repair/Modification Sheet	1	X
■ Dummy Plug		
■ Lock Collar		
■ Anti-Static Shipping Bag	1	X
■ Hard Plastic Protective End Cap	1	X
■ Pigtail with Lock Collar		
■ Spare Parts Card		
■ Dummy Plug Switch		
■ Compact Disc	1	X
■ Test Cable		
■ ECO to SBE Patch Cable		
■ White Saddle		

Checked by: CMH

Comments: _____