

SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0075
CALIBRATION DATE: 04-Nov-09

GliderAPL TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

ITS-90 COEFFICIENTS

g = 4.39945409e-003
h = 6.42679289e-004
i = 2.72238905e-005
j = 3.08686541e-006
f0 = 1000.0

IPTS-68 COEFFICIENTS

a = 3.64763522e-003
b = 5.90033476e-004
c = 1.59106496e-005
d = 3.08850198e-006
f0 = 3402.504

BATH TEMP (ITS-90)	INSTRUMENT FREQ (Hz)	INST TEMP (ITS-90)	RESIDUAL (ITS-90)
1.0000	3402.504	1.0000	-0.00002
4.5000	3678.935	4.5000	0.00003
15.0000	4605.580	15.0000	0.00002
18.5000	4948.317	18.4999	-0.00008
24.0000	5522.636	24.0001	0.00006
29.0000	6083.642	29.0000	0.00001
32.5000	6498.958	32.5000	-0.00002

Temperature ITS-90 = $1 / \{g + h[\ln(f_0/f)] + i[\ln^2(f_0/f)] + j[\ln^3(f_0/f)]\} - 273.15$ (°C)

Temperature IPTS-68 = $1 / \{a + b[\ln(f_0/f)] + c[\ln^2(f_0/f)] + d[\ln^3(f_0/f)]\} - 273.15$ (°C)

Following the recommendation of JPOTS: T_{68} is assumed to be $1.00024 * T_{90}$ (-2 to 35 °C)

Residual = instrument temperature - bath temperature

Date, Offset(mdeg C)

