



Sea-Bird Scientific  
 13431 NE 20<sup>th</sup> Street  
 Bellevue, WA 98005  
 USA

+1 425-643-9866  
 seabird@seabird.com  
 www.seabird.com

SENSOR SERIAL NUMBER: 0131  
 CALIBRATION DATE: 14-Aug-20

Glider APL CONDUCTIVITY CALIBRATION DATA  
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.01075183e+001      CPcor = -9.5700e-008 (nominal)  
 h = 1.13564166e+000      CTcor = 3.2500e-006 (nominal)  
 i = -2.61970338e-003  
 j = 2.86278481e-004

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (kHz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2.99028	0.00000	0.00000
1.0000	34.8771	2.98067	5.94280	2.98069	0.00002
4.5000	34.8575	3.28823	6.16698	3.28820	-0.00002
15.0000	34.8145	4.27142	6.83380	4.27141	-0.00001
18.4999	34.8055	4.61708	7.05297	4.61708	0.00000
24.0000	34.7956	5.17589	7.39322	5.17590	0.00001
28.9999	34.7902	5.69851	7.69746	5.69852	0.00000
32.5000	34.7873	6.07150	7.90722	6.07150	-0.00001

f = Instrument Output (kHz)

t = temperature (°C); p = pressure (decibars);  $\delta$  = CTcor;  $\epsilon$  = CPcor;

$$\text{Conductivity (S/m)} = (g + h * f^2 + i * f^3 + j * f^4) / 10 (1 + \delta * t + \epsilon * p)$$

Residual (Siemens/meter) = instrument conductivity - bath conductivity

