



Sea-Bird Scientific
 13431 NE 20th Street
 Bellevue, WA 98005
 USA

+1 425-643-9866
 seabird@seabird.com
 www.seabird.com

SENSOR SERIAL NUMBER: 0141
 CALIBRATION DATE: 06-Feb-24

Glider APL CONDUCTIVITY CALIBRATION DATA
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -9.85716580e+000
 h = 1.13080240e+000
 i = 9.34752979e-004
 j = 4.34024267e-005

CPcor = -9.5700e-008 (nominal)
 CTcor = 3.2500e-006 (nominal)

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (kHz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2.94837	0.00000	0.00000
1.0000	34.5828	2.95790	5.88717	2.95779	-0.00011
4.5000	34.5628	3.26315	6.11006	3.26322	0.00006
15.0000	34.5203	4.23913	6.77283	4.23939	0.00026
18.5000	34.5109	4.58221	6.99046	4.58215	-0.00006
24.0000	34.5001	5.13676	7.32854	5.13645	-0.00032
29.0000	34.4924	5.65521	7.63126	5.65537	0.00016
32.5000	34.4854	6.02478	7.83867	6.02324	-0.00154

f = Instrument Output (kHz)

t = temperature (°C); p = pressure (decibars); δ = CTcor; ϵ = CPcor;

$$\text{Conductivity (S/m)} = (g + h * f^2 + i * f^3 + j * f^4) / 10 (1 + \delta * t + \epsilon * p)$$

$$\text{Residual (Siemens/meter)} = \text{instrument conductivity} - \text{bath conductivity}$$

