



Sea-Bird Scientific
 13431 NE 20th Street
 Bellevue, WA 98005
 USA

+1 425-643-9866
 seabird@seabird.com
 www.seabird.com

SENSOR SERIAL NUMBER: 0141
 CALIBRATION DATE: 07-Dec-18

Glider APL CONDUCTIVITY CALIBRATION DATA
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -9.90089377e+000
 h = 1.14483973e+000
 i = -2.27285447e-003
 j = 2.59233358e-004

CPcor = -9.5700e-008 (nominal)
 CTcor = 3.2500e-006 (nominal)

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (kHz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2.94652	0.00000	0.00000
1.0000	34.7322	2.96946	5.89231	2.96946	0.00000
4.5000	34.7122	3.27587	6.11557	3.27587	-0.00000
15.0000	34.6698	4.25554	6.77952	4.25554	0.00000
18.5000	34.6607	4.59995	6.99770	4.59994	-0.00001
24.0000	34.6504	5.15667	7.33639	5.15669	0.00002
29.0000	34.6447	5.67737	7.63922	5.67736	-0.00001
32.5000	34.6407	6.04882	7.84801	6.04896	0.00014

f = Instrument Output (kHz)

t = temperature (°C); p = pressure (decibars); δ = CTcor; ϵ = CPcor;

$$\text{Conductivity (S/m)} = (g + h * f^2 + i * f^3 + j * f^4) / 10 (1 + \delta * t + \epsilon * p)$$

$$\text{Residual (Siemens/meter)} = \text{instrument conductivity} - \text{bath conductivity}$$

