



Sea-Bird Scientific  
 13431 NE 20<sup>th</sup> Street  
 Bellevue, WA 98005  
 USA

+1 425-643-9866  
 seabird@seabird.com  
 www.seabird.com

SENSOR SERIAL NUMBER: 0141  
 CALIBRATION DATE: 14-Feb-24

Glider APL CONDUCTIVITY CALIBRATION DATA  
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.00259394e+001  
 h = 1.15497614e+000  
 i = -1.09989576e-003  
 j = 1.70943882e-004

CPcor = -9.5700e-008 (nominal)  
 CTcor = 3.2500e-006 (nominal)

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (kHz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2.94854	0.00000	0.00000
1.0000	34.6276	2.96137	5.85984	2.96135	-0.00002
4.5000	34.6081	3.26701	6.08106	3.26702	0.00001
15.0000	34.5659	4.24413	6.73910	4.24417	0.00004
18.5000	34.5566	4.58762	6.95540	4.58763	0.00001
24.0000	34.5459	5.14283	7.29122	5.14276	-0.00007
29.0000	34.5389	5.66197	7.59164	5.66201	0.00003
32.5001	34.5335	6.03224	7.79848	6.03197	-0.00026

f = Instrument Output (kHz)

t = temperature (°C); p = pressure (decibars);  $\delta$  = CTcor;  $\epsilon$  = CPcor;

$$\text{Conductivity (S/m)} = (g + h * f^2 + i * f^3 + j * f^4) / 10 (1 + \delta * t + \epsilon * p)$$

Residual (Siemens/meter) = instrument conductivity - bath conductivity

