



Sea-Bird Scientific
 13431 NE 20th Street
 Bellevue, WA 98005
 USA

+1 425-643-9866
 seabird@seabird.com
 www.seabird.com

SENSOR SERIAL NUMBER: 3327
 CALIBRATION DATE: 13-Sep-24

SBE 43F OXYGEN CALIBRATION DATA

COEFFICIENTS:
 Soc = 3.6339e-004
 Foffset = -836.79
 Tau20 = 1.09
 A = -2.6856e-003
 B = 1.2407e-004
 C = -2.3958e-006
 E nominal = 0.036

NOMINAL DYNAMIC COEFFICIENTS
 D1 = 1.92634e-4
 D2 = -4.64803e-2
 H1 = -3.300000e-2
 H2 = 5.00000e+3
 H3 = 1.45000e+3

BATH OX (ml/l)	BATH TEMP (° C)	BATH SAL (PSU)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT OXYGEN (ml/l)	RESIDUAL (ml/l)
1.13	6.00	0.00	1197.735	1.13	-0.00
1.13	2.00	0.00	1159.271	1.13	-0.00
1.13	12.00	0.00	1257.531	1.13	0.00
1.14	20.00	0.00	1340.820	1.14	0.00
1.14	30.00	0.00	1451.305	1.14	0.00
1.14	26.00	0.00	1406.547	1.14	0.00
3.88	26.00	0.00	2773.147	3.88	0.00
3.88	2.00	0.00	1945.730	3.88	-0.00
3.88	30.00	0.00	2928.189	3.88	0.00
3.89	6.00	0.00	2079.423	3.89	-0.00
3.89	20.00	0.00	2557.311	3.89	-0.00
3.89	12.00	0.00	2281.762	3.89	-0.00
6.66	2.00	0.00	2738.911	6.66	0.00
6.67	6.00	0.00	2970.591	6.67	0.00
6.70	12.00	0.00	3329.164	6.70	0.00
6.71	26.00	0.00	4183.372	6.71	-0.00
6.73	30.00	0.00	4457.460	6.73	-0.00
6.74	20.00	0.00	3822.818	6.74	0.00

F = instrument output (Hz); T = temperature (°C); S = salinity (PSU); K = temperature (°K)

Oxsol(T,S) = oxygen saturation (ml/l); P = pressure (dbar)

Oxygen (ml/l) = Soc * (F + Foffset) * (1.0 + A * T + B * T² + C * T³) * Oxsol(T,S) * exp(E * P / K)

Residual (ml/l) = instrument oxygen - bath oxygen

