

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0075
CALIBRATION DATE: 14-Dec-11

GliderAPL CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

GHIJ COEFFICIENTS

g = -1.01079070e+001
h = 1.14750286e+000
i = -2.10769655e-003
j = 2.45147428e-004
CPcor = -9.5700e-008 (nominal)
CTcor = 3.2500e-006 (nominal)

ABCDM COEFFICIENTS

a = 6.02676088e-006
b = 1.14129945e+000
c = -1.00889893e+001
d = -8.70532053e-005
m = 5.2
CPcor = -9.5700e-008 (nominal)

| BATH TEMP (ITS-90) | BATH SAL (PSU) | BATH COND (Siemens/m) | INST FREQ (kHz) | INST COND (Siemens/m) | RESIDUAL (Siemens/m) |
|-----------------------|-------------------|--------------------------|--------------------|--------------------------|-------------------------|
| 22.0000 | 0.0000 | 0.00000 | 2.97325 | 0.00000 | 0.00000 |
| 1.0000 | 34.9392 | 2.98547 | 5.91137 | 2.98547 | 0.00000 |
| 4.5000 | 34.9188 | 3.29344 | 6.13445 | 3.29344 | 0.00000 |
| 14.9999 | 34.8744 | 4.27798 | 6.79793 | 4.27796 | -0.00002 |
| 18.4999 | 34.8645 | 4.62406 | 7.01602 | 4.62406 | 0.00000 |
| 23.9999 | 34.8529 | 5.18346 | 7.35456 | 5.18347 | 0.00001 |
| 29.0000 | 34.8441 | 5.70636 | 7.65716 | 5.70637 | 0.00001 |
| 32.5001 | 34.8370 | 6.07920 | 7.86561 | 6.07919 | -0.00001 |


Conductivity = $(g + hf^2 + if^3 + jf^4) / 10(1 + \delta t + \epsilon p)$ Siemens/meter

Conductivity = $(af^m + bf^2 + c + dt) / [10(1 + \epsilon p)]$ Siemens/meter

t = temperature[°C]; p = pressure[decibars]; δ = CTcor; ϵ = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients

Date, Slope Correction



● 17-Mar-11 0.9998428
▲ 14-Dec-11 1.0000000

**POST CRUISE
CALIBRATION**

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0075
CALIBRATION DATE: 14-Dec-11

GliderAPL TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

ITS-90 COEFFICIENTS

g = 4.39800857e-003
h = 6.39648787e-004
i = 2.52347891e-005
j = 2.66228490e-006
f0 = 1000.0

IPTS-68 COEFFICIENTS

a = 3.64763748e-003
b = 5.89961276e-004
c = 1.54803464e-005
d = 2.66382195e-006
f0 = 3402.851

| BATH TEMP (ITS-90) | INSTRUMENT FREQ (Hz) | INST TEMP (ITS-90) | RESIDUAL (ITS-90) |
|-----------------------|-------------------------|-----------------------|----------------------|
| 1.0000 | 3402.851 | 0.9998 | -0.00019 |
| 4.5000 | 3679.369 | 4.5003 | 0.00034 |
| 14.9999 | 4605.996 | 14.9997 | -0.00015 |
| 18.4999 | 4948.723 | 18.4997 | -0.00022 |
| 23.9999 | 5523.047 | 24.0000 | 0.00012 |
| 29.0000 | 6084.115 | 29.0004 | 0.00036 |
| 32.5001 | 6499.421 | 32.4998 | -0.00026 |

Temperature ITS-90 = $1/\{g + h[\ln(f_0/f)] + i[\ln^2(f_0/f)] + j[\ln^3(f_0/f)]\} - 273.15$ (°C)


Temperature IPTS-68 = $1/\{a + b[\ln(f_0/f)] + c[\ln^2(f_0/f)] + d[\ln^3(f_0/f)]\} - 273.15$ (°C)

Following the recommendation of JPOTS: T_{68} is assumed to be $1.00024 * T_{90}$ (-2 to 35 °C)

Residual = instrument temperature - bath temperature

Date, Offset(mdeg C)

● 17-Mar-11 -1.46
▲ 14-Dec-11 -0.00



**POST CRUISE
CALIBRATION**



SEA-BIRD ELECTRONICS, INC.

13431 NE 20th St. Bellevue, Washington 98005 USA

Phone: (425) 643-9866 Fax: (425) 643-9954 www.seabird.com

Temperature Calibration Report

| | | | |
|---------------|------------------------------|-----------------|--------------------------|
| Customer: | SEAGLIDER FABRICATION CENTER | | |
| Job Number: | 66944 | Date of Report: | 12/14/2011 |
| Model Number: | Glider | Serial Number: | 0075 Glider T/C Assembly |

Temperature sensors are normally calibrated 'as received', without adjustments, allowing a determination sensor drift. If the calibration identifies a problem, then a second calibration is performed after work is completed. The 'as received' calibration is not performed if the sensor is damaged or non-functional, or by customer request.

An 'as received' calibration certificate is provided, listing coefficients to convert sensor frequency to temperature. Users must choose whether the 'as received' calibration or the previous calibration better represents the sensor condition during deployment. In SEASOFT enter the chosen coefficients. The coefficient 'offset' allows a small correction for drift between calibrations (consult the SEASOFT manual). Calibration coefficients obtained after a repair apply only to subsequent data.

'AS RECEIVED CALIBRATION'

☒ Performed ☐ Not Performed

Date: 12/14/2011

Drift since last cal: +0.00196 Degrees Celsius/year

Comments:

'CALIBRATION AFTER REPAIR'

☐ Performed ☒ Not Performed

Date:

Drift since Last cal: Degrees Celsius/year

Comments:



SEA-BIRD ELECTRONICS, INC.

13431 NE 20th Street Bellevue, Washington 98005 USA

Phone: (425) 643-9866 Fax: (425) 643-9954 www.seabird.com

Conductivity Calibration Report

| | | | |
|---------------|------------------------------|-----------------|--------------------------|
| Customer: | SEAGLIDER FABRICATION CENTER | | |
| Job Number: | 66944 | Date of Report: | 12/14/2011 |
| Model Number: | Glider | Serial Number: | 0075 Glider T/C Assembly |

Conductivity sensors are normally calibrated 'as received', without cleaning or adjustments, allowing a determination of sensor drift. If the calibration identifies a problem or indicates cell cleaning is necessary, then a second calibration is performed after work is completed. The 'as received' calibration is not performed if the sensor is damaged or non-functional, or by customer request.

An 'as received' calibration certificate is provided, listing the coefficients used to convert sensor frequency to conductivity. Users must choose whether the 'as received' calibration or the previous calibration better represents the sensor condition during deployment. In SEASOFT enter the chosen coefficients. The coefficient 'slope' allows small corrections for drift between calibrations (consult the SEASOFT manual). Calibration coefficients obtained after a repair or cleaning apply only to subsequent data.

'AS RECEIVED CALIBRATION'

☒ Performed ☐ Not Performed

Date: 12/14/2011

Drift since last cal: -0.00050 PSU/month*

Comments:

'CALIBRATION AFTER CLEANING & REPLATINIZING'

☐ Performed ☒ Not Performed

Date:

Drift since Last cal: PSU/month*

Comments:

**Measured at 3.0 S/m*

Cell cleaning and electrode replatinizing tend to 'reset' the conductivity sensor to its original condition. Lack of drift in post-cleaning-calibration indicates geometric stability of the cell and electrical stability of the sensor circuit.

SBE SEA-BIRD ELECTRONICS, INC.

13431 NE 20th St. Bellevue, Washington 98005 USA



Phone: (425) 643-9866 Fax: (425) 643-9954 www.seabird.com

Service

Report

RMA Number

66944A

Customer Information:

Company SEAGLIDER FABRICATION CENTER

Date 1/11/2012

Contact Karl Kunkle

PO Number 100902

Serial Number 0075 Glider T/C Assembly

Model Number Glider

Services Requested:

1. Evaluate/Repair Instrumentation.
2. Perform Routine Calibration Service.

Problems Found:

Services Performed:

1. Performed initial diagnostic evaluation.
2. Performed "Post Cruise" calibration of the temperature & conductivity sensors.
3. Performed complete system check and full diagnostic evaluation.

Special Notes: