

# Instrument Checklist

Date:	4/2/2014	S/N:	BB2FLVMT-590	Order #	22170	

**Contents:** 

Description	Qty	Packed	
■ ECO Meter	1	X	
■ Calibration/Characterization Sheet	1	X	
■ Repair/Modification Sheet	1	X	
■ Dummy Plug			
■ Lock Collar			
■ Anti-Static Shipping Bag	1	X	
■ Hard Plastic Protective End Cap	1	X	
■ Pigtail with Lock Collar			
■ Spare Parts Card			
■ Dummy Plug Switch			
■ Compact Disc	1	X	
■ Test Cable			
■ ECO to SBE Patch Cable			
■ White Saddle			

Checked by:	СМН			
Comments:				

PO Box 518 620 Applegate St. Philomath OR 97370 wetlabs@wetlabs.com



(541) 929-5650 Fax (541) 929-5277 www.wetlabs.com

Date: 4/1/2014

**Customer: Univ of Hawaii** 

S/N# BB2FLVMT-590

Technician: dcm

Repairs and Modifications: Evaluated instrument and found instrument fully functional, conducted standard puck service.

Comments: New char sheets included.

PO Box 518 620 Applegate St. Philomath, OR 97370



(541) 929-5650 Fax (541) 929-5277 www.wetlabs.com

## **Scattering Meter Calibration Sheet**

3/31/2014

Wavelength: 700

S/N BB2FLVMT-590

Use the following equation to obtain "scaled" output values:

 $\beta(\theta_c)$  m<sup>-1</sup> sr<sup>-1</sup> = Scale Factor x (Output - Dark Counts)

• Scale Factor for 700 nm = 3.213E-06 (m<sup>-1</sup>sr<sup>-1</sup>)/counts

Output = meter reading counts

• Dark Counts = 48 counts

Instrument Resolution = 1.0 counts 3.26E-06 (m<sup>-1</sup>sr<sup>-1</sup>)

#### Definitions:

- Scale Factor: Calibration scale factor,  $\beta(\theta_c)$ /counts. Refer to User's Guide for derivation.
- Output: Measured signal output of the scattering meter.
- Dark Counts: Signal obtained by covering detector with black tape and submersing sensor in water.

Instrument Resolution: Standard deviation of 1 minute of collected data.

PO Box 518 620 Applegate St. Philomath, OR 97370



(541) 929-5650 Fax (541) 929-5277 www.wetlabs.com

### **Scattering Meter Calibration Sheet**

3/31/2014

Wavelength: 470

S/N BB2FLVMT-590

Use the following equation to obtain "scaled" output values:

 $\beta(\theta_c) \text{ m}^{-1} \text{ sr}^{-1} = \text{Scale Factor } x \text{ (Output - Dark Counts)}$ 

• Scale Factor for 470 nm = 1.586E-05 (m<sup>-1</sup>sr<sup>-1</sup>)/counts

Output = meter reading counts

• Dark Counts = 48 counts

Instrument Resolution = 1.4 counts 2.24E-05 (m<sup>-1</sup>sr<sup>-1</sup>)

#### Definitions:

- Scale Factor: Calibration scale factor,  $\beta(\theta_c)$ /counts. Refer to User's Guide for derivation.
- Output: Measured signal output of the scattering meter.
- Dark Counts: Signal obtained by covering detector with black tape and submersing sensor in water.

Instrument Resolution: Standard deviation of 1 minute of collected data.

PO Box 518 620 Applegate St. Philomath, OR 97370



(541) 929-5650 Fax (541) 929-5277 www.wetlabs.com

### **ECO** Chlorophyll Fluorometer Characterization Sheet

Date: 4/1/2014 S/N: BB2FLVMT-590

Chlorophyll concentration expressed in µg/l can be derived using the equation:

CHL (µg/I) = Scale Factor \* (Output - Dark counts)

Digital

Dark counts Scale Factor (SF)
Maximum Output
Resolution

50 counts 0.0155 μg/l/count 4120 counts 1.3 counts

Ambient temperature during characterization

21.0 °C

Dark Counts: Signal output of the meter in clean water with black tape over detector.

**SF**: Determined using the following equation:  $SF = x \div (output - dark counts)$ , where x is the concentration of the solution used during instrument characterization. SF is used to derive instrument output concentration from the raw signal output of the fluorometer.

Maximum Output: Maximum signal output the fluorometer is capable of.

Resolution: Standard deviation of 1 minute of collected data.

The relationship between fluorescence and chlorophyll-a concentrations in-situ is highly variable. The scale factor listed on this document was determined using a mono-culture of phytoplankton (Thalassiosira weissflogii). The population was assumed to be reasonably healthy and the concentration was determined by using the absorption method. To accurately determine chlorophyll concentration using a fluorometer, you must perform secondary measurements on the populations of interest. This is typically done using extraction-based measurement techniques on discrete samples. For additional information on determining chlorophyll concentration see "Standard Methods for the Examination of Water and Wastewater" part 10200 H, published jointly by the American Public Health Association, American Water Works Association, and the Water Environment Federation.

BB2FLVMT-590.xls Revision S 10/4/07