

Rec

10/31/2018

RBR

Conductivity Calibration Certificate

RBRmaestro C.T.D.DOrinko.FI.FI.PAR.trans.Tu.V|fast6 s/n: 80331
References: Autosal8400B#66289, MS-315#15506, SSW P160, RC#002


Reference Resistance (ohm)	Reference Conductivity (mS/cm)	Voltage Ratio, V	Measured Conductivity (mS/cm)	Calibration Error (mS/cm)	Coefficients
open	0.0000	-0.000091	0.0008	0.0008	C0: 15.059239E-3
694.020	6.2227	0.039507	6.2226	-0.0000	C1: 157.12569
331.920	13.0111	0.082711	13.0110	-0.0001	X0: 202.57342E-6
150.010	28.7891	0.183123	28.7884	-0.0007	X1: -19.139177E-6
100.014	43.1805	0.274718	43.1803	-0.0002	X2: 600E-9
75.017	57.5694	0.366290	57.5686	-0.0008	X3: 15.014443
55.514	77.7939	0.495017	77.7949	0.0009	X4: 10

Bath	Voltage Ratio	Temperature (ITS-90)	Salinity (PSS-78)	Conductivity (mS/cm)
T15S35	0.2731870	15.01444	35.0073	42.9398
T25S35	0.3371009	24.93753	34.9885	52.9903


Cell Constant @T15S35 = 4.31865 1/cm

$$C_{cor} = \frac{C_0 + C_1 * V - X_0 * (T - X_3)}{1 + X_1 * (T - X_3) + X_2 * (P - X_4)}$$

Calibration error vs. Conductivity



Calibration error vs. Temperature



Calibration Date: 26/Oct/2018
Issue Date: 26/Oct/2018
File Name: 080331_20181026_1358C.rsk

Operator: I. Shkvetz
ishkvorets


Approver: T. Akiweth
takuetteh

RBR Temperature Calibration Certificate

Logger ID: RBRmaestro Serial No: 80331 Channel No: 2

Reference Temperature, ITS-90	Voltage ratio, V	Measured Temperature, ITS-90	Calibration error	Coefficients
-4.97497	0.800532	-4.97499	-0.00002	C0: 0.003370972352645
0.66964	0.749200	0.66968	0.00004	C1: -0.000254187230509
6.89566	0.685806	6.89560	-0.00006	C2: 0.000002339522291
12.80377	0.620999	12.80394	0.00016	C3: -0.000000072199298
18.57318	0.555653	18.57293	-0.00025	
24.31703	0.490894	24.31713	0.00010	
30.11157	0.428032	30.11165	0.00008	
34.99726	0.378238	34.99720	-0.00006	

Residuals vs. Temperature



Calibration Date: 23/Oct/2018
 Issue Date: 24/Oct/2018
 Calibration ID: 28778

Operator: *Christine Mazerolle*
 cmazerolle

Approver: *I. Shkvetsov*
 ishkvetsov



Pressure Calibration Certificate

RBRmaestro C.T.D.DOrinko.FI.FI.PAR.trans.Tu.V|fast6 s/n: 80331

Sensor rating: 500 dbar s/n: I003548

Nominal accuracy: 0.05%FS (0.25 dbar)



Reference instrument: Mensor CPC6050 s/n: 41000CAM

Applied pressure, P _{app} (dbar)	Voltage ratio, V	Measured pressure, P _{meas} (dbar)	Calibration error (dbar)	Coefficients
10.1184	-0.033030	10.1199	0.0015	C0: 85.82749301
59.9995	-0.011137	59.9960	-0.0035	C1: 2283.61152354
109.9990	0.010761	110.0013	0.0023	C2: 123.26731144
159.9980	0.032604	159.9988	0.0008	C3: 9.71945717
210.0010	0.054396	209.9996	-0.0014	X0 (Patm): 10.117
259.9990	0.076136	259.9970	-0.0020	X1: -0.19150664
309.9990	0.097828	310.0022	0.0032	X2: -0.00030181
360.0000	0.119465	360.0012	0.0012	X3: 0.00000312
409.9980	0.141050	409.9969	-0.0011	X4: -0.00013445
459.9990	0.162586	459.9965	-0.0025	X5 (Tcal): 22.6
510.0000	0.184073	510.0016	0.0016	

$$P_{meas} = C_0 + C_1 \cdot V + C_2 \cdot V^2 + C_3 \cdot V^3$$

$$P_{cor} = X_0 + \frac{P_{meas} \cdot X_0 \cdot X_1 (T - X_5) - X_2 (T - X_5)^2 - X_3 (T - X_5)^3}{1 + X_4 (T - X_5)}$$

Head (mm) = 414



Calibration Date: 24/Oct/2018
 Issue Date: 24/Oct/2018
 File Name: 080331_20181024_1230P.rsk

Operator:
 dluong

Approver:
 ishkvoets



DO Calibration Certificate

RBRmaestro s/n: 80331

Sensor: Rinko-III ARO-CAV s/n: 282

Channel: 4

Reference DO (%)	Rinko DO (%)	Coefficients
0.000	3.1364788	C0: -3.241
99.842	103.9391157	C1: 1.058
		X0: -44.68289
		X1: 138.76861
		X2: -0.31466
		X3: 0.01023
		X4: 0.00400
		X5: 0.00005
		X6: 4.16000
		X7: 0.00000
		X8: 0.00000

Calibration conditions

Temperature: 24.3 °C
Patm: 101.17 kPa
Salinity: 0.0 PSU

Calibration Date: 24/Oct/2018

Issue Date: 24/Oct/2018

Operator:

T. Akwethel
takuetteh

Approver:

I. Ishkvorets
ishkvorets